




ELSEVIER

Available online at
ScienceDirect
 www.sciencedirect.com

Elsevier Masson France
EM|consulte
 www.em-consulte.com



In-house 3D printing: Why, when, and how? Overview of the national French good practice guidelines for in-house 3D-printing in maxillo-facial surgery, stomatology, and oral surgery

Roman Hossein Khonsari^{1,*}, Jeremy Adam², Mourad Benassarou³, Hélios Bertin⁴, Benjamin Billotet⁵, Jebrane Bouaoud³, Pierre Bouletreau⁶, Rachid Garmi⁷, Timothée Gellée⁸, Pierre Haen⁹, Serge Ketoff¹⁰, Géraldine Lescaille⁸, Aurélien Louvrier¹¹, Jean-Christophe Lutz¹², Masrour Makaremi¹³, Romain Nicot¹⁴, Nathalie Pham-Dang¹⁵, Morgan Praud⁴, Françoise Saint-Pierre¹⁶, Thomas Schouman³, Ludovic Sicard¹⁷, François Simon¹⁸, Thomas Wojcik¹⁴, Christophe Meyer¹¹, The French Society of Stomatology, Maxillo-Facial Surgery and Oral Surgery (SFSCMFCO)^{1,2,3,4,5,6,7,8,9,10,11,12,13,14,15,16,17,18}

¹ Service de chirurgie maxillofaciale et chirurgie plastique, Hôpital Universitaire Necker - Enfants Malades, Assistance Publique – Hôpitaux de Paris; Faculté de médecine, Université de Paris; Paris, France

² BONE 3D, Paris, France

³ Service de chirurgie maxillofaciale et stomatologie, Hôpital Universitaire Pitié-Salpêtrière, Assistance Publique – Hôpitaux de Paris, Faculté de médecine, Sorbonne Université; Paris, France

⁴ Service de chirurgie maxillofaciale et stomatologie, Centre Hospitalier Universitaire Hôtel-Dieu; Faculté de médecine, Université de Nantes; Nantes, France

⁵ ENNOIA, Besançon, France

⁶ Service de chirurgie maxillofaciale, stomatologie, chirurgie orale et chirurgie plastique de la face, Centre Hospitalier Lyon Sud, Hospices Civils de Lyon; Faculté de Médecine, Université Claude Bernard Lyon I; Lyon, France

⁷ Service de chirurgie maxillofaciale, plastique et reconstructrice, chirurgie orale et implantologie, Centre Hospitalier Universitaire Caen Normandie; Université de Caen Normandie; Caen, France

⁸ Service de chirurgie maxillofaciale et stomatologie, Unité de chirurgie orale, Hôpital Universitaire Pitié-Salpêtrière, Assistance Publique – Hôpitaux de Paris; Faculté de médecine, Sorbonne Université; Paris, France

⁹ Service de chirurgie maxillofaciale, Hôpital d'Instruction des Armées Laveran; Marseille, France

¹⁰ Service de chirurgie maxillofaciale, Groupe Hospitalier Paris Saint-Joseph, Paris, France

¹¹ Service de chirurgie maxillofaciale, stomatologie et odontologie, Centre Hospitalier Régional Universitaire de Besançon; Faculté de Médecine, Université de Franche-Comté; Besançon, France

¹² Service de chirurgie maxillofaciale et stomatologie, Centre Hospitalier Universitaire de Strasbourg; Faculté de Médecine, Université de Strasbourg; Strasbourg, France

¹³ Département d'orthopédie dento-faciale, UFR des sciences odontologiques, Bordeaux, France

¹⁴ Service de chirurgie maxillofaciale et stomatologie, Centre Hospitalier Régional Universitaire de Lille; Faculté de Médecine Henri Warembourg, Université de Lille; Lille, France

¹⁵ Service de chirurgie maxillofaciale et chirurgie plastique, Centre Hospitalier Universitaire de Clermont-Ferrand; Faculté de Médecine, Université de Clermont Auvergne; Clermont-Ferrand, France

¹⁶ Méthodologie, Sorbonne Université; Paris, France

¹⁷ Service de chirurgie orale, Hôpital Bretonneau, Assistance Publique – Hôpitaux de Paris; Faculté d'odontologie, Université de Paris; Paris, France

¹⁸ Service de d'otorhinolaryngologie et chirurgie cervico-faciale pédiatrique, Hôpital Universitaire Necker – Enfants Malades, Assistance Publique – Hôpitaux de Paris; Faculté de médecine, Université de Paris; Paris, France

ARTICLE INFO

Article History:

Received 7 August 2021

Accepted 11 August 2021

Available online xxx

Keywords:

Printing

ABSTRACT

3D-printing is part of the daily practice of maxillo-facial surgeons, stomatologists and oral surgeons. To date, no French health center is producing in-house medical devices according to the new European standards. Based on all the evidence-based data available, a group of experts from the French Society of Stomatology, Maxillo-Facial Surgery and Oral Surgery (Société Française de Chirurgie Maxillofaciale, Stomatologie et Chirurgie Orale, SFSCMFCO), provide good practice guidelines for in-house 3D-printing in maxillo-facial surgery, stomatology, and oral surgery. Briefly, technical considerations related to printers and CAD software, which

* Corresponding author at: Service de chirurgie maxillofaciale et chirurgie plastique, Hôpital Universitaire Necker – Enfants Malades, Assistance Publique – Hôpitaux de Paris; Faculté de médecine, Université de Paris; Paris, France

E-mail address: roman.khonsari@aphp.fr (R.H. Khonsari).

Three-dimensional
Maxillofacial surgery
Medical devices
Recommendations
Guidelines

were the main challenges in the last ten years, are now nearly trivial questions. The central current issues when planning the implementation of an in-house 3D-printing platform are economic and regulatory. Successful in-house 3D platforms rely on close collaborations between health professionals and engineers, backed by regulatory and logistic specialists. Several large-scale academic projects across France will soon provide definitive answers to governance and economical questions related to the use of in-house 3D printing.

© 2021 Published by Elsevier Masson SAS.

1. Introduction

3D-printing is part of the daily practice of maxillo-facial surgeons, stomatologists and oral surgeons. Additive manufacturing allows the production of custom-made medical devices (such as, among others, cutting guides for orthognathic surgery), anatomical models and various types of surgical simulators [1]. Thanks to the recent development of accessible computer-assisted design (CAD) software and the decrease in the price of professional 3D-printers, in-house 3D platforms are currently being settled in a large number of hospitals. Despite these encouraging initiatives, in the vast majority of cases, surgeons using 3D-printing still rely on external services provided by specialized companies. No French health center is currently producing in-house medical devices according to the new European standards [2] (regulation 2017/745, effective since May 2021, <https://eur-lex.europa.eu/legal-content/EN/TXT/PDF/?uri=CELEX:32017R0745&from=EN>).

A group of experts from the French Society of Stomatology, Maxillo-Facial Surgery and Oral Surgery (*Société Française de Chirurgie Maxillofaciale, Stomatologie et Chirurgie Orale*, SFSCMFCO), led by Dr Roman Hossein Khonsari (Paris, France) under the supervision of Pr. Christophe Meyer (Besançon, France) has brought together all the evidence-based data available on the in-house use of 3D-printing in 2021. Based on the results of this workgroup, we provide expert recommendations answering the following questions:

- What are the technical and logistic challenges faced when implementing in-house 3D-printing?
- What are the regulatory issues raised by in-house 3D-printing, in particular regarding the production of medical devices?
- What are the current indications of additive manufacturing in maxillo-facial surgery, stomatology and oral surgery? What are the expected clinical benefits?
- What are the economic questions related to the use of in-house 3D-printing?

The detailed analysis of the literature and the relevant references are provided in the full version of the guidelines [3]. The current article provides an outline of the main conclusions of the larger scale full-text recommendations of our Society.

2. Technical and logistic issues related to in-house 3D-printing: the first steps

Implementing in-house 3D-printing requires the availability of high-resolution medical imaging and professional software and hardware, managed by trained technicians. In order to minimize the risk of error during design, manufacture and post-processing, every step of the procedure has to be submitted to adapted risk management and quality control, including taking into account long-term maintenance.

Beyond technical considerations, the in-house implementation of a 3D-printing platform implies local organizational changes, notably in term of staff training, recruitment, and logistic circuits. The success of an in-house platform relies on the close collaboration between imaging experts, biomedical engineers, pharmacists, and hospital

management, with the constant inclusion of users (surgeons and other medical professionals) into strategic discussions [4].

Fused Deposition Modeling techniques (FDM) can be recommended for the production of standard anatomical models (for instance lower jaw or midface) [5–8]. Stereolithography (SLA) or polyjet can be recommended when more precision is required, and for the production of medical devices (surgical guides, splints) [9–11]. Importantly, printing accuracy and the mechanical properties of the produced devices can be influenced by printing parameters, the quality of post-processing, time between production and use, and even storage conditions. These fields are currently being scientifically investigated and may lead to changes in 3D-printing strategies in the future.

3. Regulatory issues: the main limitation to initiating in-house 3D-printing implementation

In the European Union, the 3D-printing of medical devices (MD) falls under the EU Regulation 2017/745, effective in France since May 2021 (<https://eur-lex.europa.eu/legal-content/EN/TXT/PDF/?uri=CELEX:32017R0745&from=EN>) [2]. Importantly, anatomical models used for patient care (for instance for discussing surgical approaches) are considered as medical devices. In brief, regulation 2017/745 implies that hospitals wishing to 3D-print medical devices need to design a risk management plan, provide proofs for pre-clinical evaluation of the devices and account for clinical follow-up. Every step of the process has to rely on CE-marked products, including software. In France, the in-house production of medical devices has to be declared to the National Security Agency of Medicines and Health Products (*Agence Nationale de Sécurité du Médicament et des Produits de Santé*, ANSM), an administrative body part of the Ministry of Health (*Ministère des Solidarités et de la Santé*), who evaluates the conformity of the production process and the quality of the 'safety measures (materio-vigilance)'. Regulatory considerations are a crucial step when building a 3D-printing platform aiming at producing medical devices; the current 2017/745 regulation is intricate enough to require systematic collaboration with regulatory specialists. To date, no health center in France has managed to match the requirement of 2017/745. While the safety concerns motivating such regulations are legitimate, they also lead to absurd situations impairing patient care. For instance, splints for orthognathic surgery are usually manufactured within maxillofacial surgery department by dental technicians. Dental technicians are unfortunately progressively disappearing from French academic maxillofacial surgery units due to economical restrictions in state-owned hospitals. A natural evolution of this traditional but endangered in-house manufacture would be the use of 3D-printing, a safer, faster and more precise technique for splint production. Nevertheless, due to the requirements of 2017/745, the in-house production of splints using 3D printers is currently very complex, and most departments rely on externalized manufacture, an expensive option poorly adapted to the practical requirements of orthognathic surgery.

4. Current recommended use of in-house 3D-printing

A large number of studies recommend the use of 3D-printed anatomical models for teaching surgery and training surgeons, as an

alternative to traditional bedside methods or as an addition to them. 3D-models can also be used for communication with patients, and for scientific exchanges between health professionals [12–17].

In clinical practice, 3D-printed models can be used for osteosynthesis plate conformation during all types of surgery, in order to increase precision and decrease the duration of procedures.

More precisely, in trauma surgery, plate conformation on 3D-printed models can be recommended in order to decrease surgery time and improve results. In this case, the model is considered as a medical device. Furthermore, and particularly for mandibular reconstruction using free fibula transfers, cutting guides can be recommended in order to reduce surgical durations and most probably improve morphological results [18–23]. Specific studies evaluating similar questions but focusing on in-house 3D-printing rather than externalized 3D-printing have led to similar conclusions [24–29].

In orthognathic surgery, 3D-printed models for plate conformation, 3D-printed splints, cutting guides and 3D-printed plates all currently benefit from sufficient precision to be recommended in standard clinical use, according to local availabilities [29]. These personalized surgery approaches all improve precision and lower surgical times [19]. As previously mentioned, the benefits of personalized approaches in orthognathic surgery have been confirmed in the specific case of in-house 3D-printing [30–32].

For dental implant placement, 3D-printed placement guides printed using SLA can be recommended as an alternative to traditional approaches, with a preference for dental-borne guides [33–35].

In brief, in-house 3D-printed custom-made medical devices for personalized surgery can be currently recommended in trauma surgery, lower jaw reconstruction using free fibula transfers and dental implant placement. Further studies are published on a daily basis assessing the use of 3D-printed medical devices in other sub-specialties of our field, such as for instance in craniostylosis surgery, free bone transfers other than fibula and orbital surgery – all with promising results – paving the way for a generalized use of personalized approaches for the surgery of craniofacial bones and jaws [3].

5. In-house 3D-printing: financial concerns

Due to the recent changes in EU regulations, it is currently difficult to precisely assess the costs related to the settlement of an in-house 3D-printing platform allowed to produce medical devices. Nevertheless, printing anatomical models and surgical guides can already be recommended based on the international literature, and lead to potential savings, mostly by reducing surgery durations [3].

When building an in-house 3D-printing platform project, economic feasibility relies on potential logistic and organizational costs, on the initial investment in printers, software and supplies, on the need for hiring new staff (maintenance, engineers) but also on the regulatory process required for producing medical devices [36–38].

In brief, in-house printing has been proven to be a source of savings for the production of anatomical models, but the question is on hold for the production of medical devices. Current exploratory projects across the country (for instance at Assistance Publique – Hôpitaux de Paris and Besançon University Hospital) will provide evidence-based answers to these issues in the near future.

6. Conclusion

In-house 3D printing will be part of standard practice in the near future. Currently, most of surgeons using additive manufacturing for producing anatomical models, medical devices and simulators still rely on externalized services. Technical considerations related to printers and CAD software, which were the main challenges in the last ten years, are now nearly trivial questions. The central current issues when planning the implementation of an in-house 3D-printing

platform are economic and regulatory. Successful in-house 3D platforms rely on close collaborations between health professionals and engineers, backed by regulatory and logistic specialists. Several large-scale academic projects across France such as in Paris and Besançon will soon provide definitive answers to governance and economical questions related to the use of in-house 3D printing.

Funding source

This article did not receive any specific grant from funding agencies in the public, commercial, or not-for-profit sectors.

Disclosure of interest

The authors report no conflicts of interest.

Acknowledgement

The Authors would acknowledge all members of the French Society of Stomatology Maxillofacial Surgery and Oral Surgery (SFSCMFCO) for their permanent support.

References

- [1] Khorsandi D, Fahimpour A, Abasian P, Saber SS, Seyedi M, Ghanavati S, et al. 3D and 4D printing in dentistry and maxillofacial surgery: Printing techniques, materials, and applications. *Acta Biomater* 2021;122:26–49. doi: [10.1016/j.actbio.2020.12.044](https://doi.org/10.1016/j.actbio.2020.12.044).
- [2] Radley-Gardner O, Beale H, Zimmermann R, editors. *Fundamental Texts On European Private Law*. editors Hart Publishing; 2016. doi: [10.5040/9781782258674](https://doi.org/10.5040/9781782258674).
- [3] French society of stomatology, maxillofacial surgery and oral surgery (SFSCMFCO). Internalized (non-industrial) 3D printing in maxillofacial surgery, stomatology and oral surgery — Recommendations for Good Practice May 2021. [Impression 3D internalisée (non industrielle) en chirurgie maxillo-faciale, stomatologie et chirurgie orale Recommendations de Bonne Pratique Mai 2021] 2021. <https://sfscmfco.fr/sfs/wp-content/uploads/2021/08/Reco-SFSCMFCO-2021-impresion-3D-medicale-internalisee.pdf> (accessed August 7, 2021).
- [4] Louvrier A, Marty P, Barrabé A, Euvrard E, Chatelain B, Weber E, et al. How useful is 3D printing in maxillofacial surgery? *J Stomatol Oral Maxillofac Surg* 2017;118:206–12. doi: [10.1016/j.jormas.2017.07.002](https://doi.org/10.1016/j.jormas.2017.07.002).
- [5] Meglioli M, Naveau A, Macaluso GM, Catros S. 3D printed bone models in oral and cranio-maxillofacial surgery: a systematic review. *3D Print Med* 2020;6:30. doi: [10.1186/s41205-020-00082-5](https://doi.org/10.1186/s41205-020-00082-5).
- [6] Msallem B, Sharma N, Cao S, Halbeisen FS, Zeilhofer H-F, Thieringer FM. Evaluation of the Dimensional Accuracy of 3D-Printed Anatomical Mandibular Models Using FFF, SLA, SLS, MJ, and BJ Printing Technology. *J Clin Med* 2020 9:E817. doi: [10.3390/jcm9030817](https://doi.org/10.3390/jcm9030817).
- [7] Hatz CR, Msallem B, Aghlmandi S, Brantner P, Thieringer FM. Can an entry-level 3D printer create high-quality anatomical models? Accuracy assessment of mandibular models printed by a desktop 3D printer and a professional device. *Int J Oral Maxillofac Surg* 2020;49:143–8. doi: [10.1016/j.ijom.2019.03.962](https://doi.org/10.1016/j.ijom.2019.03.962).
- [8] Zeller A-N, Neuhaus M-T, Fresenborg S, Zimmerer RM, Jehn P, Spalthoff S, et al. Accurate and cost-effective mandibular biomodels: a standardized evaluation of 3D-Printing via fused layer deposition modeling on soluble support structures. *J Stomatol Oral Maxillofac Surg* 2020 S2468-7855(20)30230-5. doi: [10.1016/j.jormas.2020.09.018](https://doi.org/10.1016/j.jormas.2020.09.018).
- [9] Etemad-Shahidi Y, Qallandar OB, Evenden J, Alifui-Segbaya F, Ahmed KE. Accuracy of 3-Dimensionally Printed Full-Arch Dental Models: A Systematic Review. *J Clin Med* 2020;9:E3357. doi: [10.3390/jcm9103357](https://doi.org/10.3390/jcm9103357).
- [10] Chen L, Lin W-S, Polido WD, Eckert GJ, Morton D. Accuracy, reproducibility, and dimensional stability of additively manufactured surgical templates. *J Prosthet Dent* 2019;122:309–14. doi: [10.1016/j.prosdent.2019.02.007](https://doi.org/10.1016/j.prosdent.2019.02.007).
- [11] Kim T, Lee S, Kim GB, Hong D, Kwon J, Park J-W, et al. Accuracy of a simplified 3D-printed implant surgical guide. *J Prosthet Dent* 2020;124:195–201.e2. doi: [10.1016/j.prosdent.2019.06.006](https://doi.org/10.1016/j.prosdent.2019.06.006).
- [12] Langridge B, Momin S, Coumbe B, Woin E, Griffin M, Butler P. Systematic Review of the Use of 3-Dimensional Printing in Surgical Teaching and Assessment. *J Surg Educ* 2018;75:209–21. doi: [10.1016/j.jsurg.2017.06.033](https://doi.org/10.1016/j.jsurg.2017.06.033).
- [13] Ye Z, Dun A, Jiang H, Nie C, Zhao S, Wang T, et al. The role of 3D printed models in the teaching of human anatomy: a systematic review and meta-analysis. *BMC Med Educ* 2020;20:335. doi: [10.1186/s12909-020-02242-x](https://doi.org/10.1186/s12909-020-02242-x).
- [14] Grall P, Ferri J, Nicot R. Surgical training 2.0: A systematic approach reviewing the literature focusing on oral maxillofacial surgery - Part I. *J Stomatol Oral Maxillofac Surg* 2021 S2468-7855(21)00007-0. doi: [10.1016/j.jormas.2021.01.006](https://doi.org/10.1016/j.jormas.2021.01.006).
- [15] Nicot R, Druelle C, Schlund M, Roland-Billecart T, Gwénaél R, Ferri J, et al. Use of 3D printed models in student education of craniofacial traumas. *Dent Traumatol* 2019;35:296–9. doi: [10.1111/edt.12479](https://doi.org/10.1111/edt.12479).

- [16] AlAli AB, Griffin MF, Calonge WM, Butler PE. Evaluating the Use of Cleft Lip and Palate 3D-Printed Models as a Teaching Aid. *J Surg Educ* 2018;75:200–8. doi: [10.1016/j.j Surg.2017.07.023](https://doi.org/10.1016/j.j Surg.2017.07.023).
- [17] Bertin H, Huon J-F, Praud M, Fauvel F, Salagnac J-M, Perrin J-P, et al. Bilateral sagittal split osteotomy training on mandibular 3-dimensional printed models for maxillofacial surgical residents. *Br J Oral Maxillofac Surg* 2020;58:953–8. doi: [10.1016/j.bjoms.2020.04.039](https://doi.org/10.1016/j.bjoms.2020.04.039).
- [18] Serrano C, van den Brink H, Pineau J, Prognon P, Martelli N. Benefits of 3D printing applications in jaw reconstruction: A systematic review and meta-analysis. *J Cranio-Maxillo-Fac Surg Off Publ Eur Assoc Cranio-Maxillo-Fac Surg* 2019;47:1387–97. doi: [10.1016/j.jcms.2019.06.008](https://doi.org/10.1016/j.jcms.2019.06.008).
- [19] Nilsson J, Hindocha N, Thor A. Time matters - Differences between computer-assisted surgery and conventional planning in cranio-maxillofacial surgery: A systematic review and meta-analysis. *J Cranio-Maxillo-Fac Surg Off Publ Eur Assoc Cranio-Maxillo-Fac Surg* 2020;48:132–40. doi: [10.1016/j.jcms.2019.11.024](https://doi.org/10.1016/j.jcms.2019.11.024).
- [20] Bosc R, Hersant B, Carloni R, Niddam J, Bouhassira J, De Kermadec H, et al. Mandibular reconstruction after cancer: an in-house approach to manufacturing cutting guides. *Int J Oral Maxillofac Surg* 2017;46:24–31. doi: [10.1016/j.ijom.2016.10.004](https://doi.org/10.1016/j.ijom.2016.10.004).
- [21] Mahendru S, Jain R, Aggarwal A, Aulakh HS, Jain A, Khazanchi RK, et al. CAD-CAM vs conventional technique for mandibular reconstruction with free fibula flap: A comparison of outcomes. *Surg Oncol* 2020;34:284–91. doi: [10.1016/j.sur- onc.2020.04.012](https://doi.org/10.1016/j.sur- onc.2020.04.012).
- [22] Numajiri T, Morita D, Yamochi R, Nakamura H, Tsujiko S, Sowa Y, et al. Does an In-House Computer-Aided Design/Computer-Aided Manufacturing Approach Contribute to Accuracy and Time Shortening in Mandibular Reconstruction? *J Cranio-fac Surg* 2020;31:1928–32. doi: [10.1097/SCS.0000000000006699](https://doi.org/10.1097/SCS.0000000000006699).
- [23] Naros A, Weise H, Tilsen F, Hoefert S, Naros G, Krimmel M, et al. Three-dimensional accuracy of mandibular reconstruction by patient-specific pre-bent reconstruction plates using an “in-house” 3D-printer. *J Cranio-Maxillo-Fac Surg Off Publ Eur Assoc Cranio-Maxillo-Fac Surg* 2018;46:1645–51. doi: [10.1016/j.jcms.2018.05.047](https://doi.org/10.1016/j.jcms.2018.05.047).
- [24] Moe J, Foss J, Herster R, Powell C, Helman J, Ward BB, et al. An In-House Computer-Aided Design and Computer-Aided Manufacturing Workflow for Maxillofacial Free Flap Reconstruction is Associated With a Low Cost and High Accuracy. *J Oral Maxillofac Surg Off J Am Assoc Oral Maxillofac Surg* 2021;79:227–36. doi: [10.1016/j.joms.2020.07.216](https://doi.org/10.1016/j.joms.2020.07.216).
- [25] King BJ, Park EP, Christensen BJ, Danrad R. On-Site 3-Dimensional Printing and Preoperative Adaptation Decrease Operative Time for Mandibular Fracture Repair. *J Oral Maxillofac Surg Off J Am Assoc Oral Maxillofac Surg* 2018;76:1950. doi: [10.1016/j.joms.2018.05.009](https://doi.org/10.1016/j.joms.2018.05.009).
- [26] Marschall J, Flint RL, Kushner GM, Alpert B, Azevedo B. In-house computer-aided design and manufacturing versus traditional open reduction and internal fixation for mandibular trauma. *Oral Surg Oral Med Oral Pathol Oral Radiol* 2020;130:e23. doi: [10.1016/j.oooo.2019.12.025](https://doi.org/10.1016/j.oooo.2019.12.025).
- [27] Williams FC, Hammer DA, Wentland TR, Kim RY. Immediate Teeth in Fibulas: Planning and Digital Workflow With Point-of-Care 3D Printing. *J Oral Maxillofac Surg Off J Am Assoc Oral Maxillofac Surg* 2020;78:1320–7. doi: [10.1016/j.joms.2020.04.006](https://doi.org/10.1016/j.joms.2020.04.006).
- [28] Sigron GR, Rüedi N, Chammartin F, Meyer S, Msallem B, Kunz C, et al. Three-dimensional analysis of isolated orbital floor fractures pre- and post-reconstruction with standard titanium meshes and “hybrid” patient-specific implants. *J Clin Med* 2020;9:E1579. doi: [10.3390/jcm9051579](https://doi.org/10.3390/jcm9051579).
- [29] Van den Bempt M, Liebrechts J, Maal T, Bergé S, Xi T. Toward a higher accuracy in orthognathic surgery by using intraoperative computer navigation, 3D surgical guides, and/or customized osteosynthesis plates: A systematic review. *J Cranio-Maxillo-Fac Surg Off Publ Eur Assoc Cranio-Maxillo-Fac Surg* 2018;46:2108–19. doi: [10.1016/j.jcms.2018.10.012](https://doi.org/10.1016/j.jcms.2018.10.012).
- [30] Narita M, Takaki T, Shibahara T, Iwamoto M, Yakushiji T, Kamio T. Utilization of desktop 3D printer-fabricated “Cost-Effective” 3D models in orthognathic surgery. *Maxillofac Plast Reconstr Surg* 2020;42:24. doi: [10.1186/s40902-020-00269-0](https://doi.org/10.1186/s40902-020-00269-0).
- [31] De Riu G, Vaira LA, Ligas E, Vaittinen V, Spano G, Salzano G, et al. New protocol for in-house management of computer assisted orthognathic surgery. *Br J Oral Maxillofac Surg* 2020;58:e265–71. doi: [10.1016/j.bjoms.2020.07.022](https://doi.org/10.1016/j.bjoms.2020.07.022).
- [32] Imai H, Fujita K, Yamashita Y, Yajima Y, Takasu H, Takeda A, et al. Accuracy of mandible-independent maxillary repositioning using pre-bent locking plates: a pilot study. *Int J Oral Maxillofac Surg* 2020;49:901–7. doi: [10.1016/j.ijom.2019.11.013](https://doi.org/10.1016/j.ijom.2019.11.013).
- [33] Naeini EN, Atashkadeh M, De Bruyn H, D’Haese J. Narrative review regarding the applicability, accuracy, and clinical outcome of flapless implant surgery with or without computer guidance. *Clin Implant Dent Relat Res* 2020;22:454–67. doi: [10.1111/cid.12901](https://doi.org/10.1111/cid.12901).
- [34] Walker-Finch K, Ucer C. Five-year survival rates for implants placed using digitally-designed static surgical guides: a systematic review. *Br J Oral Maxillofac Surg* 2020;58:268–76. doi: [10.1016/j.bjoms.2019.12.007](https://doi.org/10.1016/j.bjoms.2019.12.007).
- [35] Yogui FC, Verri FR, de Luna Gomes JM, Lemos CA, Cruz RS, Pellizzer EP. Comparison between computer-guided and freehand dental implant placement surgery: A systematic review and meta-analysis. *Int J Oral Maxillofac Surg* 2021;50:242–50. doi: [10.1016/j.ijom.2020.08.004](https://doi.org/10.1016/j.ijom.2020.08.004).
- [36] Spaas C, Lenssen O. Economic analysis of a low-cost virtual surgical planning protocol for mandibular reconstruction: a case series. *Br J Oral Maxillofac Surg* 2019;57:743–8. doi: [10.1016/j.bjoms.2019.06.017](https://doi.org/10.1016/j.bjoms.2019.06.017).
- [37] Lor LS, Massary DA, Chung SA, Brown PJ, Runyan CM. Cost Analysis for In-house versus Industry-printed Skull Models for Acute Midfacial Fractures. *Plast Reconstr Surg Glob Open* 2020;8:e2831. doi: [10.1097/GOX.0000000000002831](https://doi.org/10.1097/GOX.0000000000002831).
- [38] Abo Sharkh H, Makhoul N. In-house surgeon-led virtual surgical planning for maxillofacial reconstruction. *J Oral Maxillofac Surg Off J Am Assoc Oral Maxillofac Surg* 2020;78:651–60. doi: [10.1016/j.joms.2019.11.013](https://doi.org/10.1016/j.joms.2019.11.013).